

**Nepal Netra Jyoti Sangh
Central Office, Tripureswor,
Kathmandu, Nepal**

Amendment Notice No. 2

Date: April 09, 2024 (Chaitra 27, 2080)

Nepal Netra Jyoti Sangh, Central Office, Tripureswor, Kathmandu, Nepal has made the following amendments to the bid document for the Supply, Delivery, Installation, Testing and Commissioning of Hospital Equipment Package 1 (Binocular Loop, Portable Operating Microscope and Automated Visual Field Analyzer (medium Range) IFB No. NNJS/GOODS/NCB-1.01/2080/81 published on March 13, 2024, Wednesday (Falgun 30, 2080) in the "Sourya Dainik" Daily Newspaper.

Page iv, Invitation for Bids, bullet 5 bid document purchase end date:

The bid document purchase end date "Chaitra 29, 2080 (April 11, 2024)" **has been amended** as "Baishakh 06, 2081 (April 18, 2024)."

Page iv, Invitation for Bids, bullet 7, -submission date:

The bid submission date "Chaitra 30, 2080 (April 12, 2024)" **has been amended** as "Baishakh 07, 2081 (April 19, 2024)."

Page iv, Invitation for Bids, bullet 8 bid opening date:

The bid Opening date "Chaitra 30, 2080 (April 12, 2024)" **has been amended** as "Baishakh 07, 2081 (April 19, 2024)."

Page iv, Invitation for Bids, bullet 8, minimum bid security validity date:

The bid security validity date "(i.e. Shrawan 25, 2081 (August 09, 2024))" **has been amended** as (i.e. Shrawan 32, 2081 (August 16, 2024)).

Page 40, Bid Data Sheet, ITB Clause 21.1 minimum bid security validity date:

The bid security validity date "(i.e. Shrawan 25, 2081 (August 09, 2024))" **has been amended** as (i.e. Shrawan 32, 2081 (August 16, 2024)).

Page 41, Section II. Bid Data Sheet, ITB Clause 24.1, the deadline for bid submission:

The date for deadline for bid submission "Chaitra 30, 2080 (April 12, 2024)" **has been amended** as "Baishakh 07, 2081 (April 19, 2024)."



Page 41, Section II. Bid Data Sheet, ITB Clause 27.1, bid opening date:

The bid opening date "Chaitra 30, 2080 (April 12, 2024)" has been amended as "Baishakh 07, 2081 (April 19, 2024)."

Page 52, Section III. Evaluation and Qualification Criteria Page iv, 2.4.2 Specific Experience

The wordings "in at least one Data Center including installation and commissioning Contract" after the word subcontractor has been amended as "in at least one hospital equipment supply contract"

Page 67, Form Spec-1: Bidder's proposed Specification Form and Page 88, Technical Specification, Binocular Loop, Technical Specification, Bullet 6.2 Magnification

The wordings "2.0 to 5.0x" has been amended as "2.0 to 2.5x or more"

Page 67, Form Spec-1: Bidder's proposed Specification Form and Page 88, Technical Specification, Binocular Loop, Technical Specification, Bullet 6.3 Working Distance (mm):

The wordings "340 to 420" has been amended as "350 to 400 or better"

Page 67, Form Spec-1: Bidder's proposed Specification Form and Page 89, Technical Specification, Binocular Loop, Technical Specification, Bullet 6.4 Field of view Diameter (mm):

The wordings "50 to 70" has been amended as "50 or better"

Page 67, Form Spec-1: Bidder's proposed Specification Form and Page 89, Technical Specification, Binocular Loop, Technical Specification, Bullet 6.5 Mounting Option:

The wordings "Neo frame or headband" has been amended as "Frame or headband "

Page 67, Form Spec-1: Bidder's proposed Specification Form and Page 89, Technical Specification, Binocular Loop, Technical Specification, Bullet 6.6 Loupe:

The wordings "should have easy adjustable observation distance by rotating the lens barrel" has been amended as " should have mechanism for adjustments of observation distance."

Page 69, Form Spec-1: Bidder's proposed Specification Form and Page 92, Technical Specification, Portable Operating Microscope, Technical Specification, Bullet 8.2:

The wordings " The microscope provided should have continuous zoom manual magnification changer for smooth magnification during surgeries." has been amended as " The microscope provided should have continuous zoom manual magnification changer (at least 3 steps) for

[Handwritten signature]

[Handwritten signature]



smooth magnification during surgeries.."

Page 69, Form Spec-1: Bidder's proposed Specification Form and Page 92, Technical Specification, Portable Operating Microscope, Technical Specification, Bullet 8.3:

The wordings "The magnification offered by the microscope should range from 6x to 18x or better" **has been amended** as "The magnification offered by the microscope should range from 10x to 12x or better."

Page 70, Form Spec-1: Bidder's proposed Specification Form and Page 92, Technical Specification, Portable Operating Microscope, Technical Specification, Bullet 8.4

The wordings "The overall vision field should be $\phi 45$ mm" **has been amended** as "The overall vision field should be $\phi 30$ mm or better."

Page 70, Form Spec-1: Bidder's proposed Specification Form and Page 92, Technical Specification, Portable Operating Microscope, Technical Specification, Bullet 8.6:

The wordings "The microscope should offer diopter adjustment from minimum of -6D to +6D." **has been amended** as "The microscope should offer diopter adjustment. Bidder is advised to put the range offered.."

Page 70, Form Spec-1: Bidder's proposed Specification Form and Page 92, Technical Specification, Portable Operating Microscope, Technical Specification, Bullet 8.7:

The wordings "The microscope should have provision for pupillary adjustment from minimum of 55 to 75 mm" **has been amended** as "The microscope should have provision for pupillary adjustment from minimum of 55 to 70 mm or better."

Page 70, Form Spec-1: Bidder's proposed Specification Form and Page 93, Technical Specification, Portable Operating Microscope, Technical Specification, Bullet 8.9:

The wordings "The microscope should have at least cobalt, heat absorption and 1/2ND filters installed." **has been amended** as "The microscope should have at least cobalt, heat absorption filters installed."

Page 70, Form Spec-1: Bidder's proposed Specification Form and Page 93, Technical Specification, Portable Operating Microscope, Technical Specification, Bullet 8.11:

The wordings "The microscope offered should allow inclination of microscope head from at least 0-90 degrees for convenient operation" **has been amended** as "The microscope offered should allow inclination of microscope head for convenient operation."

Page 70, Form Spec-1: Bidder's proposed Specification Form and Page 93, Technical Specification, Portable Operating Microscope, Technical Specification, Bullet 8.12:

Handwritten signature

Handwritten signature



The wordings "The light source of the microscope should be halogen type for better brightness with maximum illumination of 80,000 lux or more." **has been amended** as "The light source of the microscope should be halogen type or LED type. The brightness should be proven to be adequate capacity."

Page 71, Form Spec-1: Bidder's proposed Specification Form and Page 95, Technical Specification, Portable Operating Microscope, Technical Specification, Bullet 11.1:

The wordings " Must submit ISO 9001, ISO 14001 and ISO 13485 or better for medical devices" **has been amended** as "Must submit ISO 13485 or other as applicable for medical devices"

Page 72, Form Spec-1: Bidder's proposed Specification Form and Page 95, Technical Specification, Portable Operating Microscope, Bullet 13 warranty

The wordings "Comprehensive warranty for 1222222 months after acceptance" **is amended** as "Comprehensive warranty for 12 months after acceptance"

Page 73, Form Spec-1: Bidder's proposed Specification Form and Page 97, Technical Specification, Automated Portable Operating Microscope, Bullet 6 System Configuration:

The wordings "Perimeter unit with inbuilt PC system, touch screen monitor and printer" **has been amended** as "Perimeter unit with inbuilt PC system, monitor and printer."

Page 73, Form Spec-1: Bidder's proposed Specification Form and Page 97, Technical Specification, Automated Portable Operating Microscope, Technical Specification, Bullet 7.1:

The wordings "The system should measure temporal field of view more than 80 degree" **has been amended** as " The system should measure temporal field of view at least 80 degree."

Page 73, Form Spec-1: Bidder's proposed Specification Form and Page 98, Technical Specification, Automated Portable Operating Microscope, Technical Specification, Bullet 7.2:

The wordings "The system should provide stimulus duration 0.1-9.9s" **has been amended** as "The system should provide stimulus duration at least 200ms"

Page 73, Form Spec-1: Bidder's proposed Specification Form and Page 98, Technical Specification, Automated Portable Operating Microscope, Technical Specification, Bullet 7.4:

The wordings "The system should have background illumination of 31.5 ASB " **has been amended** as " The system should have background illumination of 30 ASB or better."



Page 74, Form Spec-1: Bidder's proposed Specification Form and Page 98, Technical Specification, Automated Portable Operating Microscope, Technical Specification, Bullet 7.7:

The wordings "The system should have fixation control like hejil krakau blind spot monitor, digital eye monitoring, Eye preview" **has been amended** as " The system should have fixation control like hejil krakau blind spot monitor"

Page 74, Form Spec-1: Bidder's proposed Specification Form and Page 99, Technical Specification, Automated Portable Operating Microscope, Technical Specification, Bullet 7.8:

The wordings "The system should have variety of stimulus colour from white on white, red on white to blue on yellow " **has been amended** as "The system should have variety of stimulus colour from white on white, blue on yellow"

Handwritten signature

Handwritten signature

