

Nepal Netra Jyoti Sangh
Central Office, Tripureswor,
Kathmandu, Nepal

Clarification Notice No. 1

Date: April 09, 2024 (Chaitra 27, 2080)

Nepal Netra Jyoti Sangh, Central Office, Tripureswor, Kathmandu, Nepal has made the following clarifications on the queries raised by the bidders to the bid document for the Supply, Delivery, Installation, Testing and Commissioning of Hospital Equipment Package 1 (Binocular Loop, Portable Operating Microscope and Automated Visual Field Analyzer (medium Range) IFB No. NNJS/GOODS/NCB-1.01/2080/81 published on March 13, 2024, Wednesday (Falgun 30, 2080) in the "Sourya Dainik" Daily Newspaper.

Sl. No	Reference Page	Queries raised by the bidders	Purchaser's response
	2. Portable Operating Microscope		
1	Form Spec-1: Bidder's proposed Specification Form and Technical Specification, Portable Operating Microscope, Bullet 8.7	The wordings "The microscope should have provision for pupillary adjustment from minimum of 55 to 75 mm" is suggested to be amended as " The microscope should have provision for pupillary adjustment from minimum of 54 to 70 mm or better	Please refer to Addendum Notice no. 2
2	Form Spec-1: Bidder's proposed Specification Form and Technical Specification, Portable Operating Microscope, Bullet 11.1	The wordings " Must submit ISO 9001, ISO 14001 and ISO 13485 or better for medical devices" is suggested to be amended as "must submit ISO certificate or better for medical device"	Please refer to Addendum Notice no. 2
3	Form Spec-1: Bidder's proposed Specification Form and Technical Specification, Portable Operating Microscope, Bullet 11.2	The wordings " Must provide evidence for degree of protection against electric shock of at least TYPE B as per the specified requirement of the standard IEC 60601-1" is suggested to be deleted	Please refer to bid document
4	Form Spec-1: Bidder's proposed	The wordings " Comprehensive warranty for 1222222 months after acceptance" is	Please refer to Addendum Notice No. 2

h



	Specification Form and Technical Specification, Portable Operating Microscope, Bullet 13 warranty	suggested to be amended as ". Comprehensive warranty for 12 months after acceptance."	
3. Automated Visual Field Analyser			
5	Form Spec-1: Bidder's proposed Specification Form and Technical Specification, Automated Visual Field Analyzer, Bullet 6 System Configuration	The wordings "Perimeter unit with inbuilt PC system, touch screen monitor and printer" is suggested to be amended as ""Perimeter unit with external PC system, and printer"	Please refer to Addendum Notice No. 2
6	Form Spec-1: Bidder's proposed Specification Form and Technical Specification, Automated Visual Field Analyzer, Bullet 7.1	The wordings "The system should measure temporal field of view more than 80 degree " is suggested to be amended as "The system should measure temporal field of view at least 80 degree "	Please refer to Addendum Notice No. 2
7	Form Spec-1: Bidder's proposed Specification Form and Technical Specification, Automated Visual Field Analyzer, Bullet 7.4	The wordings "The system should have background illumination of 31.5 ASB " is suggested to be amended as " The system should have background illumination of 10 ASB "	Please refer to Addendum Notice No. 2
8	Form Spec-1: Bidder's proposed Specification Form and Technical Specification, Automated Visual Field Analyzer, Bullet 7.5	The wordings " The system should have Threshold test field 24-2, 30-2, 10-2" is suggested to be amended as "The system should have Threshold test field central 22, centralm30 and central 10"	Please refer to bid document
9	Form Spec-1: Bidder's proposed	The wordings " The system should have variety of stimulus color from white on white, red on	Please refer to Addendum Notice No.

Handwritten signature



	Specification Form and Technical Specification, Automated Visual Field Analyzer, Bullet 7.8	white to blue on yellow " is suggested to be amended as " The system should have stimulus color green on white	
10	Form Spec-1: Bidder's proposed Specification Form and Technical Specification, Automated Visual Field Analyzer, Bullet 7.9	The wordings "Stimulus Size I-V as per Goldman standards" is suggested to be amended as "Stimulus Size III as per Goldman standards."	Please refer to bid document
11	Technical Specification of Binocular Loop, Portable Operating Microscope and Automated Visual Field Analyzer	Technical Specification of Binocular Loop, Portable Operating Microscope and Automated Visual Field Analyzer are almost copy paste of Neitz, Inami and Frey respectively	Please refer to Addendum Notice No. 2

lc

Jan

()